



Certificate No. : L2619-180823

財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

This is to certify that

Lextar Electronics Corporation
Reliability Test Lab.

No.1, Keyi Rd., Kuan Yuan Science Park, Zhunan Township, Miaoli County 35059

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2005
Accreditation Number : 2619
Originally Accredited : August 27, 2012
Effective Period : August 27, 2018 to August 26, 2021
Accredited Scope : Testing Field, see described in the Appendix
Specific Accreditation Program : Accreditation Program for Laboratory of U.S. EPA
ENERGY STAR

Chung-Lin Wang
President, Taiwan Accreditation Foundation
Date : August 23, 2018



Certificate No. : L2619-180823

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Number : 2619

Laboratory Head : MAO, Tse-Min

18.01 Commodity
Lamps (Solid State Lighting Lamps-Directional)
O008 Output
IES LM-79-08
Lumen: 10 lm to 260 klm
Sample Size:
Lighting product surface < 0.234 m²
max. length <1.25 m

Approval Signatory: LO, Alice

18.01 Commodity
LED Package, Array or Module
O025 Light Source Life, Lumen Maintenance
IES LM-80-15
Lumen: 10 lm to 250 klm
Wavelength: 350 nm to 1000 nm
Intensity: 0.002 mcd to 8 kcd
Sample size: diameter 90 mm (max.)

Approval Signatory: LEE, William; LIN, Yi-Cheng; CHIU, Chaoming

18.01 Commodity
LED package
O995 LED optical measurement
CIE-127-2007
Lumen: 0.05 mlm to 250 klm
Wavelength: 360 nm to 830 nm
Intensity: 0.002 mcd to 8000 cd

Approval Signatory: LO, Alice



Certificate No. : L2619-180823

財團法人全國認證基金會
Taiwan Accreditation Foundation

Accreditation Program for Laboratory of U.S. EPA ENERGY STAR

■ 18.01 Commodity
Lamps (Solid State Lighting Lamps-Directional)
O008 Output
IES LM-79-08
Lumen: 10 lm to 260 klm
Sample Size:
Lighting product surface < 0.234 m²
max. length <1.25 m

Approval Signatory: LO, Alice

■ 18.01 Commodity
LED Package, Array or Module
O025 Light Source Life, Lumen Maintenance
IES LM-80-15
Lumen: 10 lm to 250 klm
Wavelength: 350 nm to 1000 nm
Intensity: 0.002 mcd to 8 kcd
Sample size: diameter 90 mm (max.)

Approval Signatory: LEE, William; LIN, Yi-Cheng; CHIU, Chaoming

(Null below)